

*Local Standard Weights.*

Denomination.	Amount of error tolerated in excess.
	Half the amount tolerated in deficiency.
<b>Avoirdupois Weights :—</b>	
Above 56 lb. ..	5 grains.
" 28 , and not exceeding 56 lb. ..	4 grains.
" 14 " "	3 grains.
" 2 " "	2 grains.
" 1 " "	1 grain.
" 8 oz. "	0·5 grain.
" 2 "	0·2 grain.
" 8 drams "	0·1 grain.
" 1 dram "	0·05 grain.
$\frac{1}{2}$ dram and not exceeding 1 dram ..	0·02 grain.
<b>Troy Weights :—</b>	
Above 200 oz. ..	3 grains.
" 50 , and not exceeding 200 oz. ..	2 grains.
" 30 " "	1 grain.
" 10 " "	0·5 grain.
" 2 " "	0·2 grain.
" 0·5 " "	0·1 grain.
" 0·1 " "	0·05 grain.
" 0·01 " "	0·02 grain.
" 0·005 " "	0·01 grain.
" 0·002 " "	0·005 grain.
Not exceeding 0·002 oz. ..	0·003 grain.
<b>Grain Weights :—</b>	
Above 1,000 grains.. ..	0·2 grain.
" 240 , and not exceeding 1,000 grains ..	0·1 grain.
" 40 " "	0·05 grain.
" 5 " "	0·02 grain.
" 3 " "	0·01 grain.
" 1 grain "	0·005 grain.
" 0·8 " "	0·003 grain.
" 0·05 " "	0·001 grain.
Not exceeding 0·05 grain ..	0·0003 grain.
<b>Apothecaries Weights :—</b>	
Above 2 oz. ..	0·2 grain.
" 4 drachms and not exceeding 2 oz. ..	0·1 grain.
" 2 scruples "	0·05 grain.
" 5 grains "	0·02 grain.
" 3 " "	0·01 grain.
" 1 grain "	0·005 grain.
Not exceeding 1 grain ..	0·003 grain.

**II.—METRIC WEIGHTS AND MEASURES.***Local Standard Measures of Length.*

Denomination.	Amount of error tolerated in excess or in deficiency.
Above Double Metre or 2 metres ..	2·5 millimetres.
" Metre and not exceeding Double Metre or 2 metres ..	0·5 millimetre.
" Decimetre " " Metre or 1,000 millimetres ..	0·25 millimetre.
" Centimetre " " Decimetre or 0·1 metre.. ..	0·1 millimetre.
" Millimetre " " Centimetre or 0·01 metre ..	0·05 millimetre.
Not exceeding Millimetre or 0·001 metre ..	0·01 millimetre.