

Local Standard Measures of Capacity.

Denomination.	Amount of error tolerated in excess or in deficiency.
Grammes weight of water as measured by a graduated glass tube.	
Above 10 litres (Dekalitre) ..	10 grammes.
" 5 " and not exceeding 10 litres (Dekalitre) ..	5 grammes.
" 2 " " 5 "	2·5 grammes.
" 1 litre " 2 "	1 gramme.
" 0·1 litre (Decilitre) and not exceeding 1 litre ..	0·5 grammes.
" 0·02 " " 0·1 " (Decilitre)	0·3 grammes.
" 0·005 " " 0·02 "	0·2 grammes.
" 0·002 " " 0·005 "	0·1 grammes.
Not exceeding 0·002 litre ..	0·05 grammes.

Local Standard Cubic Measures.

Denomination.	Amount of error tolerated in excess or in deficiency.
Grammes weight of water as measured by a graduated glass tube.	
Above 100 cubic centimetres ..	0·5 grammes.
" 20 " " and not exceeding 100 cubic centimetres ..	0·3 grammes.
" 5 " " " 20 "	0·2 grammes.
" 2 " " " 5 "	0·1 grammes.
Not exceeding 2 cubic centimetres ..	0·05 grammes.

Local Standard Weights.

Denomination.	Amount of error tolerated in excess.
	Half the amount tolerated in deficiency.
Above 5 kilograms ..	200 milligrams.
" 1 kilogram and not exceeding 5 kilograms ..	180 milligrams.
" 200 grammes " " 1 kilogram ..	30 milligrams.
" 50 " " " 200 grammes ..	18 milligrams.
" 10 " " " 50 "	6 milligrams.
" 2 " " " 10 "	3 milligrams.
" 5 decigrams " " 2 "	1 milligram.
" 2 " " " 5 decigrams ..	0·7 milligram.
" 5 centigrams " " 2 "	0·3 milligram.
" 2 " " " 5 centigrams ..	0·2 milligram.
" 5 milligrams " " 2 "	0·06 milligram.
Not exceeding 5 milligrams ..	0·02 milligram.

III.—LOCAL STANDARD GAS MEASURES.

Denomination.	Amount of error tolerated.
Gasholders ..	1 or 0·25 per centum <i>in excess or in deficiency.</i>
Test gas-meters ..	1 or 0·25 per centum <i>fast or slow.</i>